

Project	Job Number	Set	Sample Id	PCB's	DCBP
631-23019	2450511	1	BW-5	<0.05	96.0 %
631-23019	2450511	1	BE-11	<0.05	98.0 %
631-23019	2450511	1	DTW-0122	<0.05	101. %
631-23019	2450511	1	Spiked BE-11	103. %	97.0 %
631-23019	2450511	1	Blank	<0.05	98.0 %
631-23019	2450511	1	QC Standard (found)	92.0%	101. %
631-23019	2450511	1	QC Standard (expected)	100. %	100. %
631-23019	2450511	1	LCL	65.0%	65.0 %
631-23019	2450511	1	UCL	135. %	135. %
631-23019	2450511	1	Repeat BW-5	<0.05	94.0 %